

UV HANDHELD LAMP MATRIX

SPECTRO-UV®

Part Number	ASTM E3022	ISO 3059	RRES:90061	AIRBUS AITM6	PRATT & WHITNEY	Nominal Steady-State UV-A Intensity at 15 in (38 cm)	UV-A Coverage at 15 in (38 cm) at minimum 1,200 $\mu\text{W}/\text{cm}^2$	Power Source	Run Time	White Light	Applications
TRITAN SERIES											
TRI-365SBLC	X	X	X	X	X	4,500 $\mu\text{W}/\text{cm}^2$	5.5 in (14 cm) Ø	AC Cord	—	Y	Aerospace, Industrial, NDT, Forensics
TRI-365HC	X	X				8,000 $\mu\text{W}/\text{cm}^2$	6 in (15 cm) Ø	AC Cord	—	Y	Aerospace, Industrial, NDT, Forensics
TRI-365HB						9,000 $\mu\text{W}/\text{cm}^2$	6.5 in (16.5 cm) Ø	AC Cord	—	Y	Aerospace, Industrial, NDT, Forensics
TRI-365DB						4,750 $\mu\text{W}/\text{cm}^2$	5.5 in (14 cm) Ø	AC Cord	—	Y	Aerospace, Industrial, NDT, Forensics
TRI-365MDB						1,400 $\mu\text{W}/\text{cm}^2$	5.5 in (14 cm) Ø	AC Cord or Li-Ion Battery	7 hrs	Y	Aerospace, Industrial, NDT, Forensics
TRI-365MHB						9,000 $\mu\text{W}/\text{cm}^2$	6.5 in (16.5 cm) Ø	AC Cord or Li-Ion Battery	7 hrs	Y	Aerospace, Industrial, NDT, Forensics

Key Selling Points (Tritan):

- Built-in fans for continuous use
- Thermal cut-off protection
- Replaceable cord options
- LED protection covers



UV HANDHELD LAMP MATRIX

Technical specs for UV-A lamps



Pratt & Whitney

PARAMETERS	ASTM E3022-18	ISO 3059:2012	RRES 90061:2014	AITM6-1001 Issue 15	UV-A Inspection Requirements
Wavelength (nm)	360 to 370	365 ± 5	365 ± 5	365 ± 5	365 peak
FMHM (nm)	≤ 15	≤ 30	≤ 20 @ 50% OP ≤ 30 @ 10% OP	Per ASTM E3022	N/A
Intensity (μW/cm ²) at 15"	≥ 2,000	Min: 1,000 Max: 5,000	Min: 1,200 Max: 5,000	Min: 1,200 Max: 5,000	Min: 1,000 Max: 5,000
Beam profile	≥ 5-inch at ≥ 1,000 μW/cm ²	N/A	Even irradiation	Per ASTM E3022	High uniformity
Visible light emission & Ambient visible light	≤ 21.5 lux	≤ 20 lux	≤ 5 lux at min WD ≤ 20 lux at max WD	Per ASTM E3022	≤ 20 lux
Electronic control Current ripple: Battery shut off: Temp cut off:	Ripple < 5% Intensity < 1,000 N/A	N/A	Ripple < 6% Current drop > 3% Over temp condition	Per ASTM E3022	Thermal management EMC
Report	Each lamp + Lamp model	N/A	Unit test Validation report	Per ASTM E3022	Per ASTM E3022

SPECTRO-UV®

www.Spectro-UV.com

4 Dubon Ct., Farmingdale, NY 11735

866-230-7305